

**Attachment 4**

**Clinical trial report of SARS-CoV-2 Antigen Test Kit  
(Colloidal Gold Chromatographic Immunoassay)**

Version: A01

Issue Date: 2020.10.30

## **Explanation**

1. Medical institutions in charge of clinical trials should conduct clinical trials in a fair and objective manner in accordance with clinical trial protocols in a serious and responsible manner, and write clinical trial reports.
2. This trial must be conducted by a clinical trial institution with an experienced attending physician or more, or a person with intermediate job title or above who has the relevant work experience.
3. This clinical trial category is clinical validation, and this report should be written with reference to the Technical Guidelines for Clinical Trials of In Vitro Diagnostic Reagents.

Principal researchers of this test

Clinical tester	Position/Title	Responsible for testing content	Current unit
RenLiZhang	Chief technician	Principal researcher, responsible for the organization of clinical tests, review of clinical test protocols and reports.	Shenzhen Center for Diseases Control and Prevention
YinSun	technician	Researcher	Shenzhen Center for Diseases Control and Prevention
YaLanHuang	technician	Researcher	Shenzhen Center for Diseases Control and Prevention
DeZhongXie	technician	Head of statistics	Shenzhen Microprofit Biotech Co., Ltd.

## **1. Clinical validation time**

This clinical evaluation is conducted from Mar 2020 to Oct 20th, 2020.

## **2. Background information for clinical evaluation**

The 2019 novel coronavirus (2019-nCoV) was discovered around the end of 2019 due to the emergence of novel viral pneumonia cases of unknown cause and was officially named by the World Health Organization (SARS-CoV-2) as COVID-19 on January 12, 2020. Coronaviruses are a large family of viruses that are known to cause colds and more serious diseases such as the Middle East Respiratory Syndrome (MERS) and Severe Acute Respiratory Syndrome (SARS). The 2019 novel coronavirus, however, is a new coronavirus strain that has not previously seen in humans.

Common signs of a person infected with coronavirus include respiratory symptoms, fever, cough, shortness of breath and difficulty breathing. In more serious cases, infection can cause pneumonia, severe acute respiratory syndrome, kidney failure, and even death.

The SARS-CoV-2 Antigen Test Kit (Colloidal Gold Chromatographic Immunoassay) Test developed by our company can help diagnose whether patients are infected with the Novel Coronavirus. It has further enriched the detection methods of Novel Coronavirus, expanded the supply of detection reagents, and fully served the needs of epidemic prevention and control.

## **3. Test purposes**

The SARS-CoV-2 Antigen Test Kit (Colloidal Gold Chromatographic Immunoassay) produced by Shenzhen Microprofit Biotech Co., Ltd. is used to verify the feasibility of clinical evaluation and the reliability of test results. The purpose of research of the clinical test is to calculate the consistency percentage of negative/positive and the total consistency percentage and Kappa coefficient by statistically analyzing test results through comparative experimental research.

## **4. Test design**

### **4.1 Test plan selection and reasons**

In vitro diagnostic reagents for testing and reference reagents are used to conduct comparative research tests on clinically suspected Novel Coronavirus Nasopharyngeal swab samples, and it is proved that the in vitro diagnostic reagents used in the test can achieve the expected assistance in infection of the Novel Coronavirus.

### **4.2 Sample volume required**

The total number of clinical trials of this product is not less than 100 cases. The samples is classified into the positive group and the negative group as per the test results of the reference product. Meanwhile, the samples shall be tested via the qualitative test strip tested and by reference product from the same patient and then the test results of the product

tested and the reference product shall be compared, with statistical analysis being made.

#### 4.3 Sample inclusion/exclusion certification.

The positive group and negative group in this experiment are applicable to the following inclusion/exclusion criteria

Positive group inclusion: Meets 2 of the following 3 criteria, it is inclusion into positive sample group:

Antibody test is positive.

PCR Test is positive.

CT test results and symptoms are clinically positive.

Positive group exclusion:

Samples only Meet 1 of the 3 criteria of inclusion; it is exclusion out of positive sample group. Negative inclusion:

Meets 3 of the following 4 criteria, it is inclusion into negative sample group:

Antibody test is negative.

PCR test is negative.

CT test results and symptoms are clinically negative.

No history of novel coronavirus exposure within 14 day.

Negative exclusion:

Any sample that does not meet the inclusion criteria is excluded out of the negative group.

#### 4.4 Sample collection, processing, and storage

Sample collection: It is applicable to the diagnosis of the coronavirus from the samples of Nasopharyngeal swabs or Nasopharyngeal aspirates. Use freshly collected samples for optimal test performance. Inadequate sample collection or improper sample handling may yield a false-negative result.

Completely insert the sterilized swab supplied in this kit into the Nasopharyngeal basin and swab several times to collect the epidermal cells of the mucus. It is recommended to collect sample from Nasopharyngeal basin for more accurate results.

In vitro diagnostic reagents and reference products for testing

Test in vitro diagnostic reagents

Name: SARS-CoV-2 Antigen Test Kit (Colloidal Gold Chromatographic Immunoassay)

Specification: 25 tests/kit

LOT: 20200701

Expiry: July, 2021 (Tentative)

Storage Conditions: Store in a dry place at 2-30°C, protected from light. After opening the inner package, the test card will become invalid due to moisture absorption. Please use it within 1 hour.

Source: Shenzhen Microprofit Biotech Co., Ltd.

Reference products

Real-Time Fluorescent RT-PCR Kit for Detecting SARS-2019- nCoV from BGI Genomics Co. Ltd., which had obtained by the EUA.

Storage Conditions: Store in a dry place at 2-8°C, protected from light.

## 5. Experiment method

### 5.1. Sample collection

Nasopharyngeal swab collection method:

1. Tip the patient's head back and collect sample from the nostril that has more mucus (head should be inclined from vertical for proper specimen collection).
2. Insert the swab through the nostril entry and then slowly move along the bottom of the Nasopharyngeal cavity (Move gently to avoid traumatic bleeding).
3. When the tip of the swab reaches the posterior wall of the nasopharyngeal cavity, gently rotate it several times. (Collect as much secretion as possible)
4. To prevent reflex coughing, stop for one minute.
5. Slowly remove the swab.
6. Test the sample as soon as possible

Nasal swabs collection method:

1. Wash or disinfect your hands, ensuring they are dry before testing
2. Carefully remove the sterile nasal swab from the packaging, ensuring you do not touch the end of the cotton swab
3. Gently insert the cotton swab end of the nasal swab into the left nostril to a depth of 2.5 cm (1 inch)
4. Rotate the nasal swab 5 times against the walls (mucous membrane) of the left nostril to ensure adequate sampling
5. Remove the nasal swab and insert it into the right nostril to a depth of 2.5cm (1 inch)
6. Rotate the nasal swab 5 times against the walls (mucous membrane) of the right nostril to ensure adequate sampling
7. Ensure you use the same nasal swab for both nostrils to ensure adequate sampling

### 5.2. Sample treatment

#### 5.2.1 Swab samples:

- Prepare a sample treatment tube and open its cap, add about 20 drops (500 $\mu$ L) sample treatment solution to the treatment tube.
- Place the swab sample in the tube, dip the swab in the solution and rotate it 10 times, then break the swab at the swab node, and leave the lower half in the treatment tube.
- Cover treatment tube cap, shake the tube upside down, then wait for 5 minutes of reaction.

#### 5.2.2 Viral Transport Media

Note: Confirm whether VTM is suitable for the kit.

- Prepare a sample treatment tube and open its cap, add about 8 drops (200 $\mu$ L) sample treatment solution to the treatment tube.
- Take 200 $\mu$ L sample(VTM)to the tube
- Cover treatment tube cap, shake the tube upside down, then wait for 5 minutes of reaction.

## 6. Statistical methods of statistical analysis of clinical research data

### A. Methods evaluating clinical performance

The product tested shall be subject to tests through the sample of different types, with statistics on the results. Meanwhile, different types of sample of the subjects shall be subject to determination by the product tested synchronously, and then the determination results of both shall be compared. The test results recorded shall be subject to statistical analysis upon completion of determination of all clinical samples, to calculate the consistency percentage of negative/positive and the total consistency percentage. Afterwards, equivalence of both shall be evaluated as per

these statistical indexes.

### B. Statistical method

The products launched on the market shall be subject to comparative study and evaluation. Kappa inspection: each sample shall be tested with the product tested and the reference product respectively, and then the consistency in statistical results of these two inspection methods shall be compared through Kappa inspection.

The data shall be subject to Kappa inspection and analysis and the Kappa coefficient shall be calculated. Favorable consistency can be proven if Kappa is >0.8. The consistency in test results of the product tested and the reference product is evaluated as per the evaluation standards.

## 7. Standards of clinical evaluation

The coincidence rate shall be calculated by comparing with the reference product whose marketing is approved. The product performance shall meet the following requirements.

Coincidence rate of negative: the sample whose test results are negative for both the product tested and the reference product and the proportion in the sample whose test results are negative for the reference product shall be more than 95%.

Coincidence rate of positive: the sample whose test results are positive for both the product tested and the reference product and the proportion in the sample whose test results are positive for the reference product shall be more than 85%.

Total coincidence rate: the sample whose test results are the same for the product tested and the reference product and its proportion in the total number of samples shall be more than 90%.

If the coincidence rate of positive/negative can meet clinical requirements, two methods or Products are considered as equivalent; If the coincidence rate of positive/negative is greatly different, the clinical scheme should be re-designed. Kappa consistency analysis shall be adopted for statistical analysis of reference reagents.

The results of the product tested are statistical materials and can be per the table below:

Method		2019-nCoV nucleic acid		Total Results
		positive	negative	
SARS-CoV-2 Antigen Test Kit (Colloidal Gold Chromatographic Immunoassay)	Result			
	positive	A	B	A+B
	negative	C	D	C+D
Total Results		A+C	B+D	A+B+C+D

$$\text{Clinical sensitivity} = A/(A+C) * 100\%$$

$$\text{Clinical specificity} = D/(B+D) * 100\%$$

$$\text{Accuracy: } (A+D)/(A+B+C+D) * 100\%$$

$$Po = (A+D)/(A+B+C+D) * 100\%$$

$$Pe = ((A+B) (A+C) + (A+B) (B+D)) / (A+B+C+D)^2$$

$$\text{Kappa: } (Po - Pe) / (1 - pe)$$

If conducting Kappa consistency analysis for the base data above, high consistency can be judged if the Kappa coefficient is >0.8, and both systems are considered as equivalent. Consistency is considered if 0.4<Kappa coefficient <0.8, and the coincidence rate of positive/negative shall be compared, with statistical analysis being made. Two such systems are considered as inconsistent and in-equivalent if the Kappa coefficient is <0.4.

Provisions for amendments to clinical validation

In general, the clinical validation should not be changed. Any modification to the project during the test should be explained, and the time, reason, process of change, and whether there is a record of the change are explained in detail and its impact on the evaluation of the entire research result is explained.

Results and Analysis of Clinical Tests

7.1 The results of Nasopharyngeal swab

In total, 668 Nasopharyngeal swab test samples are included for the unit and all test samples included are tested.

7.1. 1.Statistics on test results and those of the product tested are as follows:

Method		2019-nCoV Nucleic Acid Test Kit		Total Results
SARS-CoV-2 Antigen Test Kit (Colloidal Gold Chromatographic Immunoassay)	Results	Positive	Negative	
	Positive	342	0	342
	Negative	26	300	326
Total Results		368	300	668

Clinical sensitivity

$$=342/368=92.93\%(95\%CI:89.82\%~95.33\%)$$

$$\text{Clinical specificity} = 300/300=100\% (95\%CI:98.78\% \sim 100\%)$$

$$\text{Accuracy:}(342+300)/(342+0+26+300)*100\%=96.11\%(95\%CI:94.35\%~97.4\%)$$

$$Pe = (342*368+342*300) / (668*668) = 0.5120$$

$$\text{Kappa:} (Po - Pe) / (1 - pe) = 0.9220$$

According to the above table, 300 are proven negative of 300 negative specimens, 342 are proven positive of 368 positive specimens. The sensitivity and accuracy are more than 90%, indicating favorable consistency with the reference product. The Kappa=0.9220 > 0.8, indicating favorable and high consistency of two methods and equivalence of two such systems.

7.1.2.PCR Cycle Threshold (CT) analysis of PPA at each stage

Cycle Threshold(CT)	# of RT-PCR positive	fluorecare® SARS-CoV-2 Antigen Test (Colloidal Gold Chromatographic Immunoassay)		
		# of positive results	PPA	NPA
<20	39	39	100.00%	100%
<25	112	111	99.11%	
<30	217	214	98.62%	
<35	297	292	98.32%	
<40	368	342	92.93%	

7.2 The results of Nasal swab

In total, 424 Nasal swabs test samples are included for the unit and all test samples included are tested.



7.2. 1.Statistics on test results and those of the product tested are as follows:

Method		2019-nCoV Nucleic Acid Test Kit		Total Results
SARS-CoV-2 Antigen Test Kit (Colloidal Gold Chromatographic Immunoassay)	Results	Positive	Negative	
	Positive	117	0	117
	Negative	7	300	307
Total Results		124	300	424

Clinical sensitivity =  $117/124=94.35\%$  (88.71%~97.70%)

Clinical specificity =  $300/300=100\%$  (95%CI:98.78% ~100%)

Accuracy:  $(117+300)/(117+0+7+300)*100%=98.35\%$  (96.63%~99.33%)

$Pe = (117*124+117*300)/(424*424) = 0.2759$

Kappa:  $(Po - Pe)/(1-pe) = 0.9594$

According to the above table, 300 are proven negative of 300 negative specimens, 117 are proven positive of 124 positive specimens. The sensitivity and accuracy are more than 90%, indicating favorable consistency with the reference product. The Kappa=0.9594 > 0.8, indicating favorable and high consistency of two methods and equivalence of two such systems.

7.2.2.PCR Cycle Threshold (CT) analysis of PPA at each stage

Cycle Threshold(CT)	# of RT-PCR positive	fluorecare® SARS-CoV-2 Antigen Test (Colloidal Gold Chromatographic Immunoassay)		
		# of positive results	PPA	NPA
<20	44	44	100.00%	100%
<25	82	82	100.00%	
<30	107	107	100.00%	
<35	123	117	95.12%	
<40	124	117	94.35%	

### 8. Analysis on Inconsistency in test results

NO.	Age	Sex	Sample type	SARS-CoV-2 Antigen Test Kit (Colloidal Gold Chromatographic Immunoassay)	2019-nCoV nucleic acid test kit (RT-PCR)	Cycle Threshold(CT)	Clinical diagnostic
35	43	Female	Nasal swab	Negative	Positive	31	Infection 1 days
373	43	Female	Nasal swab	Negative	Positive	32	Infection 7 days
49	64	Male	Nasal swab	Negative	Positive	33	Infection 2 days
379	54	Male	Nasal swab	Negative	Positive	33	Infection 1 days
20	56	Female	Nasal swab	Negative	Positive	34	Infection 4 days
30	35	Female	Nasal swab	Negative	Positive	34	Infection 0 days
383	33	Male	Nasal swab	Negative	Positive	34	Infection 5 days

## **9. Discussion and Conclusions**

### 9.1 discussion

A Results of comparative analysis of the product tested and the reference product:

Test results of Swab specimen tested and the reference result: both the coincidence rate of negative/positive and the total coincidence rate are larger than 90%, indicating favorable consistency with the reference product. In the analysis results of Kappa inspection, Kappa is proven  $>0.8$ , indicating favorable and high consistency of both methods. Both systems are proven equivalent.

### 9.2 Test conclusions

By analyzing the test results of the product tested and the reference product, the consistency percentage of negative/positive and the total consistency percentage are proven high. Moreover, according to the results of statistical analysis, there is no remarkable difference in test results of both, indicating favorable consistency in diagnosis and equivalence of two such systems and can be used for auxiliary diagnosis of those suffering from pneumonia triggered by COVID-19.

## **10. Quality control methods**

### On-site quality control

During this study, clinical implementors appointed clinical inspectors to conduct regular on-site supervision visits to the research hospital. Through monitoring visits, it is found that all the contents of the research plan are strictly observed, and the correctness of the research data is also guaranteed. Participating researchers have undergone unified training, unified recording methods and judgment standards. The entire clinical trial process is conducted under strict operation, and the test content is complete and authentic. All observations and findings in the clinical trials have been verified and the data are reliable. The conclusions in the clinical trials are derived from the original data.

### Quality control of clinical experiment process

During the evaluation, quality control is performed daily to ensure that the product is under control. Strict quality control is performed for each trial to ensure the quality of clinical trials.

## **11. Prediction of adverse events**

Because SARS-CoV-2 Antigen Test Kit (Colloidal Gold Chromatographic Immunoassay) is an in vitro diagnostic reagent product, no direct contact with patients is required in clinical trials, no test report is provided to patients, and the test results are only used for comparative studies. It involves personal privacy, does not serve as a basis for auxiliary diagnosis, does not bring any risk to the subject, and does not cause adverse events.

## Annex 1

Evaluation data sheet of sample test results (sample type: Nasopharyngeal swab)

Sample Number	Age	Gender	Sample type	Diagnosis	SARS-CoV-2 Antigen Test Kit results	RT-PCR test results	Sample collection time	Cycle Thresh old(CT)	The onset of symptoms
1	50	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/3	26	2020/9/2
2	51	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/31	25	2020/8/29
3	49	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/27	15	2020/8/26
4	33	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/31	31	2020/8/25
5	55	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/16	19	2020/9/13
6	60	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/7	27	2020/9/2
7	58	Female	Nasopharyngeal swab	Symptomatic	Negative	Positive	2020/9/9	23	2020/9/8
8	64	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/8	26	2020/9/8
9	21	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/19	26	2020/9/16
10	31	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/11	26	2020/9/8
11	55	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/31	12	2020/8/26
12	67	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/25	28	2020/9/19
13	52	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/7	31	2020/9/5
14	35	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/18	27	2020/9/17
15	33	Male	Nasopharyngeal swab	Symptomatic	Negative	Positive	2020/9/14	27	2020/9/13
16	55	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/27	17	2020/8/26
17	69	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/22	14	2020/9/15
18	30	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/17	26	2020/9/13
19	47	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/20	29	2020/9/14
20	56	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/11	22	2020/9/9
21	40	Female	Nasopharyngeal swab	Symptomatic	Negative	Positive	2020/9/24	27	2020/9/20
22	40	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/17	26	2020/9/13
23	41	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/9	25	2020/9/9
24	32	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/25	31	2020/9/18
25	32	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/19	12	2020/9/13
26	45	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/4	31	2020/9/1
27	29	Female	Nasopharyngeal swab	Symptomatic	Negative	Positive	2020/9/18	32	2020/9/17
28	25	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/2	27	2020/8/28
29	65	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/24	34	2020/9/17
30	35	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/16	19	2020/9/16
31	47	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/7	30	2020/9/2
32	32	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/24	15	2020/8/24
33	32	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/3	22	2020/9/2
34	41	Female	Nasopharyngeal swab	Asymptomatic	Negative	Positive	2020/9/1	38	2020/8/29
35	43	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/25	36	2020/9/20
36	34	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/14	13	2020/9/10
37	57	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/17	34	2020/9/14
38	41	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/5	12	2020/8/30

39	63	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/1	32	2020/8/27
40	52	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/27	26	2020/8/25
41	66	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/5	21	2020/8/31
42	41	Female	Nasopharyngeal swab	Asymptomatic	Negative	Positive	2020/8/31	37	2020/8/29
43	60	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/11	28	2020/9/5
44	52	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/7	14	2020/9/4
45	27	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/9	27	2020/9/5
46	53	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/26	26	2020/9/21
47	37	Male	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/21	37	2020/9/20
48	38	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/11	23	2020/9/8
49	64	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/11	27	2020/9/5
50	43	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/16	34	2020/9/13
51	69	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/19	30	2020/9/13
52	51	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/22	26	2020/9/17
53	40	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/21	31	2020/9/19
54	60	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/25	27	2020/9/19
55	46	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/22	14	2020/9/20
56	47	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/20	22	2020/9/13
57	44	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/11	25	2020/9/8
58	29	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/28	31	2020/8/26
59	41	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/10	27	2020/9/9
60	35	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/8	19	2020/9/3
61	28	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/27	30	2020/8/24
62	68	Male	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/2	38	2020/9/2
63	44	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/12	13	2020/9/8
64	55	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/1	12	2020/8/30
65	65	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/26	32	2020/8/26
66	49	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/31	26	2020/8/26
67	31	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/16	14	2020/9/10
68	32	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/25	17	2020/9/20
69	47	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/27	11	2020/9/21
70	32	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/28	28	2020/8/28
71	48	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/23	35	2020/9/17
72	38	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/23	36	2020/9/16
73	38	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/20	21	2020/9/18
74	28	Female	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/9	37	2020/9/5
75	49	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/14	32	2020/9/12
76	32	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/21	26	2020/9/20
77	25	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/2	25	2020/8/31
78	44	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/10	26	2020/9/6
79	81	Female	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/11	37	2020/9/7
80	41	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/4	27	2020/8/31
81	48	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/2	36	2020/8/26
82	33	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/24	34	2020/9/22

83	21	Female	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/1	39	2020/8/26
84	29	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/31	19	2020/8/29
85	69	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/26	31	2020/8/26
86	72	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/29	22	2020/8/28
87	61	Female	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/22	37	2020/9/22
88	70	Female	Nasopharyngeal swab	Asymptomatic	Negative	Positive	2020/9/21	37	2020/9/20
89	56	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/4	26	2020/8/30
90	58	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/4	34	2020/9/3
91	44	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/12	18	2020/9/11
92	51	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/20	26	2020/9/14
93	28	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/17	30	2020/9/16
94	54	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/31	19	2020/8/30
95	32	Female	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/18	39	2020/9/18
96	68	Female	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/6	38	2020/8/30
97	52	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/15	25	2020/9/14
98	65	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/1	25	2020/8/30
99	73	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/27	22	2020/8/25
100	80	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/18	29	2020/9/16
101	26	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/18	32	2020/9/17
102	32	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/31	36	2020/8/30
103	70	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/26	18	2020/9/22
104	34	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/24	22	2020/9/19
105	75	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/23	29	2020/9/20
106	48	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/11	19	2020/9/4
107	52	Female	Nasopharyngeal swab	Asymptomatic	Negative	Positive	2020/9/12	39	2020/9/10
108	59	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/10	21	2020/9/4
109	36	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/3	25	2020/8/28
110	32	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/31	30	2020/8/30
111	72	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/17	18	2020/9/11
112	32	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/8	30	2020/9/6
113	62	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/29	35	2020/8/24
114	27	Female	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/7	37	2020/9/3
115	62	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/23	36	2020/9/18
116	71	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/30	26	2020/8/26
117	78	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/15	20	2020/9/8
118	28	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/1	33	2020/8/26
119	51	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/3	24	2020/8/31
120	53	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/16	28	2020/9/12
121	72	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/3	22	2020/9/2
122	33	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/12	19	2020/9/11
123	57	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/1	24	2020/8/27
124	58	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/18	20	2020/9/15
125	35	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/14	29	2020/9/13
126	37	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/31	34	2020/8/27

127	60	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/6	19	2020/9/5
128	26	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/28	23	2020/8/27
129	22	Female	Nasopharyngeal swab	Asymptomatic	Negative	Positive	2020/9/12	38	2020/9/9
130	80	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/11	31	2020/9/10
131	51	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/23	24	2020/9/22
132	72	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/4	21	2020/9/3
133	43	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/7	28	2020/9/2
134	53	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/28	31	2020/8/28
135	46	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/13	27	2020/9/8
136	47	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/10	31	2020/9/8
137	23	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/19	27	2020/9/14
138	65	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/9	35	2020/9/6
139	31	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/11	29	2020/9/6
140	56	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/4	21	2020/9/1
141	21	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/24	33	2020/9/22
142	32	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/14	32	2020/9/13
143	28	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/1	32	2020/9/1
144	53	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/17	26	2020/9/10
145	72	Female	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/12	38	2020/9/11
146	38	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/10	22	2020/9/5
147	41	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/18	23	2020/9/11
148	32	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/20	20	2020/9/15
149	74	Female	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/12	37	2020/9/10
150	62	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/23	17	2020/9/16
151	25	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/3	19	2020/8/28
152	67	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/9	23	2020/9/8
153	54	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/29	24	2020/8/26
154	31	Female	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/14	38	2020/9/10
155	31	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/6	21	2020/9/2
156	76	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/30	21	2020/8/24
157	39	Male	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/22	37	2020/9/21
158	69	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/10	32	2020/9/5
159	58	Female	Nasopharyngeal swab	Asymptomatic	Negative	Positive	2020/9/23	37	2020/9/17
160	28	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/2	31	2020/9/1
161	53	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/12	34	2020/9/9
162	71	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/18	26	2020/9/18
163	24	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/9	27	2020/9/5
164	46	Female	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/14	37	2020/9/7
165	42	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/7	28	2020/9/7
166	72	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/11	28	2020/9/7
167	30	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/9	31	2020/9/7
168	32	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/3	35	2020/8/30
169	21	Female	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/11	37	2020/9/5
170	74	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/1	25	2020/9/1

171	35	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/2	19	2020/8/29
172	67	Female	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/8/27	38	2020/8/26
173	60	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/15	25	2020/9/11
174	68	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/28	31	2020/9/21
175	51	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/18	27	2020/9/15
176	66	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/8	32	2020/9/7
177	64	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/26	36	2020/8/25
178	56	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/22	28	2020/9/21
179	58	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/6	21	2020/9/2
180	68	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/20	27	2020/9/17
181	38	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/19	19	2020/9/17
182	45	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/23	27	2020/9/16
183	40	Female	Nasopharyngeal swab	Asymptomatic	Negative	Positive	2020/9/4	39	2020/9/4
184	59	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/2	20	2020/8/28
185	35	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/6	33	2020/8/31
186	61	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/17	33	2020/9/12
187	44	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/25	28	2020/9/22
188	66	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/29	32	2020/8/25
189	56	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/1	25	2020/8/27
190	61	Male	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/23	37	2020/9/20
191	70	Female	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/19	37	2020/9/16
192	58	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/26	27	2020/9/20
193	38	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/25	29	2020/9/19
194	57	Male	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/7	38	2020/9/6
195	58	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/2	18	2020/9/1
196	49	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/25	35	2020/9/21
197	32	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/4	34	2020/9/3
198	63	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/5	27	2020/9/4
199	36	Male	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/8/30	38	2020/8/30
200	52	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/4	33	2020/9/3
201	54	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/4	25	2020/8/28
202	64	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/30	21	2020/8/24
203	46	Female	Nasopharyngeal swab	Asymptomatic	Negative	Positive	2020/9/2	37	2020/8/28
204	53	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/5	28	2020/9/4
205	55	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/26	36	2020/8/26
206	24	Male	Nasopharyngeal swab	Asymptomatic	Negative	Positive	2020/9/18	39	2020/9/16
207	24	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/8	27	2020/9/4
208	74	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/15	23	2020/9/8
209	50	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/9	23	2020/9/6
210	64	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/26	32	2020/8/24
211	69	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/16	23	2020/9/12
212	50	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/11	19	2020/9/11
213	46	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/18	27	2020/9/12
214	27	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/4	22	2020/9/2

215	68	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/14	22	2020/9/12
216	60	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/21	26	2020/9/20
217	76	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/18	20	2020/9/17
218	25	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/14	36	2020/9/11
219	41	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/15	28	2020/9/12
220	29	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/25	36	2020/9/21
221	61	Female	Nasopharyngeal swab	Asymptomatic	Negative	Positive	2020/9/14	39	2020/9/9
222	49	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/1	24	2020/8/27
223	41	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/19	30	2020/9/18
224	41	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/5	25	2020/9/4
225	70	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/2	28	2020/8/31
226	60	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/21	32	2020/9/19
227	24	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/3	36	2020/8/28
228	69	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/15	20	2020/9/12
229	77	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/26	21	2020/9/22
230	30	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/10	29	2020/9/6
231	61	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/4	34	2020/8/29
232	80	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/20	34	2020/9/15
233	31	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/28	21	2020/8/24
234	58	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/8	23	2020/9/7
235	57	Male	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/2	38	2020/9/2
236	75	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/19	27	2020/9/18
237	53	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/8	23	2020/9/5
238	58	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/14	30	2020/9/9
239	33	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/28	34	2020/8/24
240	56	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/22	32	2020/9/19
241	74	Female	Nasopharyngeal swab	Asymptomatic	Negative	Positive	2020/8/31	39	2020/8/29
242	66	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/16	22	2020/9/9
243	31	Male	Nasopharyngeal swab	Asymptomatic	Negative	Positive	2020/9/1	39	2020/9/1
244	67	Female	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/23	38	2020/9/22
245	57	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/30	33	2020/8/30
246	40	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/30	36	2020/8/27
247	79	Female	Nasopharyngeal swab	Symptomatic	Negative	Positive	2020/9/19	35	2020/9/13
248	33	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/4	21	2020/8/31
249	74	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/10	29	2020/9/9
250	44	Male	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/28	38	2020/9/21
251	57	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/17	29	2020/9/12
252	80	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/25	26	2020/9/19
253	41	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/27	29	2020/8/24
254	28	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/25	28	2020/9/22
255	28	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/30	33	2020/8/30
256	42	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/10	27	2020/9/9
257	75	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/26	20	2020/9/21
258	63	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/26	26	2020/8/25



259	22	Female	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/8/27	37	2020/8/26
260	29	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/3	26	2020/8/29
261	40	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/25	25	2020/9/22
262	78	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/7	22	2020/9/1
263	38	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/27	34	2020/9/22
264	67	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/1	28	2020/9/1
265	23	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/20	20	2020/9/15
266	75	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/21	30	2020/9/19
267	44	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/18	26	2020/9/17
268	75	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/24	30	2020/9/18
269	80	Female	Nasopharyngeal swab	Asymptomatic	Negative	Positive	2020/9/26	39	2020/9/22
270	51	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/21	19	2020/9/17
271	66	Male	Nasopharyngeal swab	Asymptomatic	Negative	Positive	2020/8/29	38	2020/8/25
272	60	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/23	26	2020/9/19
273	66	Male	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/22	38	2020/9/20
274	80	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/1	23	2020/8/26
275	31	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/5	33	2020/8/30
276	32	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/31	18	2020/8/24
277	51	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/5	21	2020/9/2
278	27	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/23	32	2020/9/17
279	45	Male	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/2	38	2020/8/29
280	59	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/15	21	2020/9/14
281	72	Male	Nasopharyngeal swab	Symptomatic	Negative	Positive	2020/9/8	36	2020/9/6
282	23	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/24	22	2020/9/18
283	67	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/1	32	2020/8/31
284	76	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/9	24	2020/9/5
285	68	Male	Nasopharyngeal swab	Asymptomatic	Negative	Positive	2020/9/22	39	2020/9/15
286	40	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/15	28	2020/9/9
287	77	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/25	22	2020/9/20
288	60	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/27	24	2020/8/27
289	66	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/2	34	2020/8/30
290	68	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/2	25	2020/8/31
291	79	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/29	25	2020/8/28
292	30	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/3	26	2020/8/31
293	80	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/11	33	2020/9/10
294	64	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/15	25	2020/9/14
295	33	Male	Nasopharyngeal swab	Symptomatic	Negative	Positive	2020/8/29	36	2020/8/27
296	27	Female	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/20	38	2020/9/16
297	59	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/5	29	2020/8/31
298	23	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/7	36	2020/9/1
299	27	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/29	26	2020/8/26
300	27	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/9	28	2020/9/3
301	31	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/14	18	2020/9/8
302	39	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/8	20	2020/9/5

303	30	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/24	24	2020/8/24
304	38	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/29	32	2020/8/27
305	36	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/15	20	2020/9/15
306	70	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/15	36	2020/9/13
307	39	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/29	17	2020/8/28
308	45	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/3	31	2020/9/2
309	36	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/24	23	2020/9/21
310	30	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/4	34	2020/8/31
311	30	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/23	26	2020/9/16
312	53	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/6	22	2020/9/4
313	57	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/16	26	2020/9/15
314	37	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/30	23	2020/8/30
315	76	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/31	31	2020/8/29
316	78	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/2	30	2020/9/1
317	32	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/5	32	2020/8/31
318	66	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/21	26	2020/9/15
319	35	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/11	29	2020/9/7
320	68	Male	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/8/26	36	2020/8/24
321	79	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/17	25	2020/9/10
322	47	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/4	31	2020/9/2
323	52	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/18	32	2020/9/16
324	52	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/2	34	2020/8/27
325	74	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/13	26	2020/9/7
326	24	Female	Nasopharyngeal swab	Asymptomatic	Negative	Positive	2020/9/11	38	2020/9/6
327	44	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/9	30	2020/9/6
328	56	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/26	23	2020/8/24
329	67	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/7	30	2020/9/7
330	26	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/22	29	2020/9/18
331	54	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/20	30	2020/9/16
332	68	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/8	23	2020/9/7
333	45	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/7	30	2020/9/4
334	53	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/3	31	2020/8/28
335	50	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/8	17	2020/9/6
336	76	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/14	34	2020/9/7
337	36	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/23	22	2020/9/16
338	66	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/28	26	2020/9/21
339	75	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/6	24	2020/9/2
340	23	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/18	20	2020/9/16
341	70	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/30	27	2020/8/28
342	35	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/15	32	2020/9/12
343	76	Male	Nasopharyngeal swab	Asymptomatic	Negative	Positive	2020/9/1	39	2020/8/31
344	50	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/8	30	2020/9/3
345	65	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/16	25	2020/9/14
346	42	Male	Nasopharyngeal swab	Asymptomatic	Negative	Positive	2020/9/2	39	2020/8/28

347	66	Female	Nasopharyngeal swab	Asymptomatic	Negative	Positive	2020/9/11	39	2020/9/6
348	49	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/26	31	2020/8/24
349	59	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/22	18	2020/9/17
350	65	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/31	31	2020/8/26
351	64	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/12	27	2020/9/11
352	61	Female	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/2	39	2020/8/26
353	27	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/5	23	2020/9/4
354	39	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/31	29	2020/8/26
355	35	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/10	32	2020/9/7
356	33	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/5	18	2020/9/3
357	30	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/27	22	2020/9/21
358	49	Female	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/15	37	2020/9/12
359	80	Female	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/16	22	2020/9/11
360	68	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/27	21	2020/8/25
361	52	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/2	27	2020/8/27
362	30	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/21	20	2020/9/15
363	63	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/2	25	2020/8/31
364	31	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/22	20	2020/9/21
365	47	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/8/30	27	2020/8/28
366	72	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/21	24	2020/9/19
367	38	Male	Nasopharyngeal swab	Symptomatic	Positive	Positive	2020/9/21	21	2020/9/20
368	70	Male	Nasopharyngeal swab	Asymptomatic	Positive	Positive	2020/9/11	38	2020/9/4
369	21	Male	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
370	20	Female	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
371	37	Female	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
372	53	Female	Nasopharyngeal swab	Symptomatic	Negative	Negative	2020/8/3	/	/
373	69	Male	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
374	43	Female	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
375	60	Male	Nasopharyngeal swab	Symptomatic	Negative	Negative	2020/8/3	/	/
376	55	Female	Nasopharyngeal swab	Symptomatic	Negative	Negative	2020/8/3	/	/
377	42	Female	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
378	47	Female	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
379	37	Female	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
380	58	Male	Nasopharyngeal swab	Symptomatic	Negative	Negative	2020/8/3	/	/
381	70	Female	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
382	33	Male	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
383	43	Male	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
384	69	Male	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
385	29	Female	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
386	26	Female	Nasopharyngeal swab	Symptomatic	Negative	Negative	2020/8/3	/	/
387	28	Female	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
388	59	Male	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
389	70	Female	Nasopharyngeal swab	Symptomatic	Negative	Negative	2020/8/3	/	/
390	69	Male	Nasopharyngeal swab	Symptomatic	Negative	Negative	2020/8/3	/	/
391	66	Male	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
392	54	Male	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
393	51	Male	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
394	60	Male	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/













650	66	Female	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
651	54	Male	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
652	56	Male	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
653	45	Male	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
654	44	Female	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
655	47	Female	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
656	49	Female	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
657	57	Female	Nasopharyngeal swab	Symptomatic	Negative	Negative	2020/8/3	/	/
658	68	Female	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
659	63	Male	Nasopharyngeal swab	Symptomatic	Negative	Negative	2020/8/3	/	/
660	25	Female	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
661	34	Female	Nasopharyngeal swab	Symptomatic	Negative	Negative	2020/8/3	/	/
662	36	Female	Nasopharyngeal swab	Symptomatic	Negative	Negative	2020/8/3	/	/
663	27	Male	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
664	28	Female	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
665	43	Male	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
666	42	Female	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
667	49	Female	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/
668	25	Female	Nasopharyngeal swab	Asymptomatic	Negative	Negative	2020/8/3	/	/

## Annex 2

## Evaluation data sheet of sample test results (sample type: Nasal swab)

Sample Number	Age	Gender	Sample type	Diagnosis	SARS-CoV-2 Antigen Test Kit results	RT-PCR test results	Sample collection time	Cycle Threshold (CT)	The onset of symptoms
1	50	Female	Nasal swab	Symptomatic	Positive	Positive	2020/4/29	16	2020/4/24
2	51	Female	Nasal swab	Symptomatic	Positive	Positive	2020/4/18	22	2020/4/15
3	49	Female	Nasal swab	Symptomatic	Positive	Positive	2020/4/15	29	2020/4/12
4	33	Female	Nasal swab	Symptomatic	Positive	Positive	2020/3/5	27	2020/3/5
5	55	Female	Nasal swab	Asymptomatic	Positive	Positive	2020/8/2	33	2020/7/26
6	60	Female	Nasal swab	Asymptomatic	Positive	Positive	2020/4/15	23	2020/4/14
7	58	Female	Nasal swab	Symptomatic	Positive	Positive	2020/5/1	18	2020/4/25
8	64	Female	Nasal swab	Symptomatic	Positive	Positive	2020/3/14	18	2020/3/9
9	21	Male	Nasal swab	Symptomatic	Positive	Positive	2020/6/25	17	2020/6/25
10	31	Female	Nasal swab	Symptomatic	Positive	Positive	2020/7/10	24	2020/7/9
11	55	Female	Nasal swab	Asymptomatic	Positive	Positive	2020/5/20	18	2020/5/17
12	67	Female	Nasal swab	Symptomatic	Positive	Positive	2020/5/16	22	2020/5/13
13	52	Male	Nasal swab	Symptomatic	Positive	Positive	2020/2/15	28	2020/2/8
14	35	Female	Nasal swab	Symptomatic	Positive	Positive	2020/6/20	27	2020/6/19
15	33	Male	Nasal swab	Symptomatic	Positive	Positive	2020/7/27	24	2020/7/24
16	55	Male	Nasal swab	Symptomatic	Positive	Positive	2020/7/5	26	2020/6/28
17	69	Male	Nasal swab	Symptomatic	Positive	Positive	2020/6/5	17	2020/6/4
18	30	Male	Nasal swab	Symptomatic	Positive	Positive	2020/2/28	22	2020/2/23
19	47	Male	Nasal swab	Symptomatic	Positive	Positive	2020/7/6	18	2020/6/29
20	56	Female	Nasal swab	Symptomatic	Negative	Positive	2020/6/5	34	2020/6/1
21	40	Female	Nasal swab	Symptomatic	Positive	Positive	2020/7/16	18	2020/7/11
22	40	Male	Nasal swab	Symptomatic	Positive	Positive	2020/2/24	18	2020/2/22
23	41	Female	Nasal swab	Asymptomatic	Positive	Positive	2020/5/6	31	2020/4/29
24	32	Female	Nasal swab	Symptomatic	Positive	Positive	2020/5/16	17	2020/5/13
25	32	Male	Nasal swab	Symptomatic	Positive	Positive	2020/3/15	24	2020/3/9
26	45	Male	Nasal swab	Symptomatic	Positive	Positive	2020/5/17	28	2020/5/13
27	29	Female	Nasal swab	Symptomatic	Positive	Positive	2020/7/24	23	2020/7/18
28	25	Female	Nasal swab	Symptomatic	Positive	Positive	2020/7/21	17	2020/7/18
29	65	Female	Nasal swab	Asymptomatic	Positive	Positive	2020/2/19	18	2020/2/13
30	35	Female	Nasal swab	Symptomatic	Negative	Positive	2020/6/30	34	2020/6/30
31	47	Female	Nasal swab	Symptomatic	Positive	Positive	2020/3/2	28	2020/2/27
32	32	Male	Nasal swab	Symptomatic	Positive	Positive	2020/4/9	23	2020/4/3
33	32	Female	Nasal swab	Symptomatic	Positive	Positive	2020/2/22	21	2020/2/20
34	41	Female	Nasal swab	Symptomatic	Positive	Positive	2020/7/31	28	2020/7/31
35	43	Female	Nasal swab	Symptomatic	Negative	Positive	2020/7/29	31	2020/7/28
36	34	Female	Nasal swab	Symptomatic	Positive	Positive	2020/4/19	22	2020/4/13
37	57	Male	Nasal swab	Symptomatic	Positive	Positive	2020/2/14	26	2020/2/13
38	41	Female	Nasal swab	Symptomatic	Positive	Positive	2020/3/10	17	2020/3/7
39	63	Female	Nasal swab	Symptomatic	Positive	Positive	2020/6/7	21	2020/6/5
40	52	Female	Nasal swab	Symptomatic	Positive	Positive	2020/2/17	21	2020/2/11
41	66	Male	Nasal swab	Symptomatic	Positive	Positive	2020/2/18	18	2020/2/17
42	41	Female	Nasal swab	Symptomatic	Positive	Positive	2020/5/31	19	2020/5/27

43	60	Male	Nasal swab	Symptomatic	Positive	Positive	2020/2/17	21	2020/2/15
44	52	Male	Nasal swab	Symptomatic	Positive	Positive	2020/5/12	18	2020/5/11
45	27	Male	Nasal swab	Asymptomatic	Positive	Positive	2020/7/30	31	2020/7/29
46	53	Female	Nasal swab	Symptomatic	Positive	Positive	2020/4/30	27	2020/4/25
47	37	Male	Nasal swab	Asymptomatic	Positive	Positive	2020/2/22	21	2020/2/20
48	38	Female	Nasal swab	Symptomatic	Positive	Positive	2020/5/18	17	2020/5/17
49	64	Male	Nasal swab	Symptomatic	Negative	Positive	2020/5/6	33	2020/5/4
50	43	Male	Nasal swab	Symptomatic	Positive	Positive	2020/1/30	17	2020/1/26
51	69	Male	Nasal swab	Symptomatic	Positive	Positive	2020/6/13	18	2020/6/7
52	21	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
53	20	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
54	37	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
55	53	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
56	69	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
57	43	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
58	60	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
59	55	Male	Nasal swab	Symptomatic	Negative	Negative	2020/9/12	/	/
60	42	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
61	47	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
62	37	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
63	58	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
64	70	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
65	33	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
66	43	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
67	69	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
68	29	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
69	26	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
70	28	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
71	59	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
72	70	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
73	69	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
74	66	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
75	54	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
76	51	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
77	60	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
78	40	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
79	52	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
80	47	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
81	62	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
82	56	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
83	35	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
84	40	Female	Nasal swab	Symptomatic	Negative	Negative	2020/9/12	/	/
85	34	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
86	22	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
87	31	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
88	37	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
89	51	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
90	60	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
91	49	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
92	30	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/

93	25	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
94	26	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
95	46	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
96	22	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
97	40	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
98	39	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
99	47	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
100	50	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
101	25	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
102	44	Female	Nasal swab	Symptomatic	Negative	Negative	2020/9/12	/	/
103	23	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
104	39	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
105	39	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
106	56	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
107	25	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
108	29	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
109	38	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
110	38	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
111	26	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
112	67	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
113	51	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
114	41	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
115	23	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
116	31	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
117	33	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
118	41	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
119	35	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
120	42	Male	Nasal swab	Symptomatic	Negative	Negative	2020/9/12	/	/
121	36	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
122	28	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
123	68	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
124	63	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
125	67	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
126	41	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
127	44	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
128	49	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
129	50	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
130	55	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
131	29	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
132	34	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
133	40	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
134	21	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
135	67	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
136	36	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
137	32	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
138	27	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/

139	28	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
140	41	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
141	65	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
142	46	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
143	56	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
144	25	Female	Nasal swab	Symptomatic	Negative	Negative	2020/9/12	/	/
145	49	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
146	70	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
147	34	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
148	63	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
149	42	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
150	20	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
151	21	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
152	37	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
153	29	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
154	53	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
155	29	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
156	57	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
157	47	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
158	50	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
159	39	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
160	49	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
161	36	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
162	42	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
163	23	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
164	55	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
165	58	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
166	33	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
167	39	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
168	21	Male	Nasal swab	Symptomatic	Negative	Negative	2020/9/12	/	/
169	32	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
170	64	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
171	58	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
172	54	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
173	32	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
174	56	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
175	21	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
176	64	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
177	49	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
178	21	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
179	48	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
180	25	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
181	47	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
182	41	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
183	45	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/

184	48	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
185	61	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
186	33	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
187	36	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
188	32	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
189	55	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
190	49	Female	Nasal swab	Symptomatic	Negative	Negative	2020/9/12	/	/
191	23	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
192	43	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
193	52	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
194	28	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
195	36	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
196	23	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
197	26	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
198	43	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
199	29	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
200	67	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
201	46	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
202	27	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
203	49	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
204	22	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
205	45	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
206	54	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
207	54	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
208	22	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
209	44	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
210	65	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
211	43	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
212	33	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
213	44	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
214	64	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
215	49	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
216	42	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
217	60	Female	Nasal swab	Symptomatic	Negative	Negative	2020/9/12	/	/
218	54	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
219	67	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
220	49	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
221	56	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
222	51	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
223	60	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
224	33	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
225	48	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
226	57	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
227	56	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
228	65	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
229	66	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
230	61	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
231	47	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
232	29	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
233	34	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/

234	47	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
235	47	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
236	41	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
237	32	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
238	41	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
239	24	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
240	61	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
241	31	Female	Nasal swab	Symptomatic	Negative	Negative	2020/9/12	/	/
242	64	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
243	31	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
244	45	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
245	59	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
246	20	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
247	55	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
248	26	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
249	51	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
250	45	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
251	57	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
252	23	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
253	22	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
254	37	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
255	37	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
256	53	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
257	28	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
258	36	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
259	42	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
260	21	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
261	44	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
262	38	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
263	39	Female	Nasal swab	Symptomatic	Negative	Negative	2020/9/12	/	/
264	50	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
265	68	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
266	46	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
267	43	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
268	40	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
269	34	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
270	50	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
271	36	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
272	40	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
273	36	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
274	70	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
275	32	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
276	33	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
277	64	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
278	67	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
279	42	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
280	42	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
281	36	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
282	54	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
283	37	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/

284	52	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
285	61	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
286	43	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
287	28	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
288	42	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
289	46	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
290	64	Female	Nasal swab	Symptomatic	Negative	Negative	2020/9/12	/	/
291	25	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
292	26	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
293	68	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
294	37	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
295	37	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
296	68	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
297	21	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
298	56	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
299	44	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
300	26	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
301	43	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
302	59	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
303	26	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
304	65	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
305	28	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
306	37	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
307	35	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
308	65	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
309	56	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
310	59	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
311	29	Female	Nasal swab	Symptomatic	Negative	Negative	2020/9/12	/	/
312	45	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
313	58	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
314	39	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
315	20	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
316	33	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
317	68	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
318	21	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
319	28	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
320	40	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
321	60	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
322	29	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
323	25	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
324	60	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
325	65	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
326	35	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
327	36	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
328	49	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
329	36	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
330	50	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
331	28	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
332	37	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
333	66	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/



334	54	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
335	56	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
336	45	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
337	44	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
338	47	Female	Nasal swab	Symptomatic	Negative	Negative	2020/9/12	/	/
339	49	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
340	57	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
341	68	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
342	63	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
343	25	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
344	34	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
345	36	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
346	27	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
347	28	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
348	43	Male	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
349	42	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
350	49	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
351	25	Female	Nasal swab	Asymptomatic	Negative	Negative	2020/9/12	/	/
352	51	Female	Nasal swab	Symptomatic	Positive	Positive	2020/7/9	15	2020/7/7
353	40	Male	Nasal swab	Symptomatic	Positive	Positive	2020/7/18	16	2020/7/12
354	60	Female	Nasal swab	Symptomatic	Positive	Positive	2020/7/9	15	2020/7/5
355	46	Male	Nasal swab	Symptomatic	Positive	Positive	2020/7/25	19	2020/7/23
356	47	Female	Nasal swab	Symptomatic	Positive	Positive	2020/7/24	17	2020/7/19
357	44	Female	Nasal swab	Symptomatic	Positive	Positive	2020/7/14	15	2020/7/12
358	29	Male	Nasal swab	Symptomatic	Positive	Positive	2020/7/9	15	2020/7/9
359	41	Female	Nasal swab	Symptomatic	Positive	Positive	2020/7/12	15	2020/7/5
360	35	Male	Nasal swab	Symptomatic	Positive	Positive	2020/7/8	15	2020/7/7
361	28	Female	Nasal swab	Symptomatic	Positive	Positive	2020/7/3	15	2020/7/2
362	68	Male	Nasal swab	Asymptomatic	Positive	Positive	2020/7/10	15	2020/7/9
363	44	Male	Nasal swab	Symptomatic	Positive	Positive	2020/7/16	16	2020/7/13
364	55	Female	Nasal swab	Asymptomatic	Positive	Positive	2020/7/28	18	2020/7/22
365	65	Male	Nasal swab	Asymptomatic	Positive	Positive	2020/7/6	15	2020/7/4
366	49	Male	Nasal swab	Symptomatic	Positive	Positive	2020/7/15	16	2020/7/14
367	31	Male	Nasal swab	Symptomatic	Positive	Positive	2020/7/18	17	2020/7/13
368	32	Female	Nasal swab	Symptomatic	Positive	Positive	2020/8/1	20	2020/7/28
369	47	Male	Nasal swab	Asymptomatic	Positive	Positive	2020/7/30	17	2020/7/24
370	47	Male	Nasal swab	Symptomatic	Positive	Positive	2020/8/8	29	2020/8/1
371	37	Male	Nasal swab	Symptomatic	Positive	Positive	2020/8/30	20	2020/8/27
372	47	Male	Nasal swab	Asymptomatic	Positive	Positive	2020/8/19	30	2020/8/14
373	43	Female	Nasal swab	Symptomatic	Negative	Positive	2020/7/31	32	2020/7/24
374	59	Female	Nasal swab	Symptomatic	Positive	Positive	2020/8/11	27	2020/8/6
375	25	Female	Nasal swab	Symptomatic	Positive	Positive	2020/8/23	18	2020/8/18
376	60	Male	Nasal swab	Symptomatic	Positive	Positive	2020/8/7	23	2020/8/2
377	26	Female	Nasal swab	Asymptomatic	Positive	Positive	2020/8/29	26	2020/8/24
378	47	Male	Nasal swab	Symptomatic	Positive	Positive	2020/8/18	21	2020/8/16
379	54	Male	Nasal swab	Symptomatic	Negative	Positive	2020/8/11	33	2020/8/10
380	50	Male	Nasal swab	Symptomatic	Positive	Positive	2020/8/12	25	2020/8/8
381	56	Male	Nasal swab	Symptomatic	Positive	Positive	2020/9/4	21	2020/8/29
382	28	Male	Nasal swab	Symptomatic	Positive	Positive	2020/9/3	30	2020/8/30
383	33	Male	Nasal swab	Asymptomatic	Negative	Positive	2020/9/5	34	2020/8/31

384	39	Female	Nasal swab	Symptomatic	Positive	Positive	2020/8/4	22	2020/7/31
385	42	Male	Nasal swab	Symptomatic	Positive	Positive	2020/8/21	17	2020/8/18
386	29	Male	Nasal swab	Symptomatic	Positive	Positive	2020/8/2	21	2020/8/1
387	70	Female	Nasal swab	Symptomatic	Positive	Positive	2020/8/21	26	2020/8/20
388	51	Female	Nasal swab	Asymptomatic	Positive	Positive	2020/9/7	19	2020/9/1
389	69	Female	Nasal swab	Symptomatic	Positive	Positive	2020/8/17	27	2020/8/15
390	26	Male	Nasal swab	Symptomatic	Positive	Positive	2020/8/7	30	2020/8/2
391	35	Female	Nasal swab	Symptomatic	Positive	Positive	2020/8/19	26	2020/8/14
392	60	Female	Nasal swab	Symptomatic	Positive	Positive	2020/8/3	24	2020/8/3
393	66	Female	Nasal swab	Symptomatic	Positive	Positive	2020/8/21	30	2020/8/16
394	40	Male	Nasal swab	Symptomatic	Positive	Positive	2020/7/29	29	2020/7/28
395	40	Male	Nasal swab	Symptomatic	Positive	Positive	2020/8/18	21	2020/8/13
396	62	Male	Nasal swab	Symptomatic	Positive	Positive	2020/8/14	24	2020/8/10
397	70	Female	Nasal swab	Asymptomatic	Positive	Positive	2020/8/5	27	2020/8/4
398	49	Female	Nasal swab	Symptomatic	Positive	Positive	2020/9/2	21	2020/8/29
399	51	Female	Nasal swab	Symptomatic	Positive	Positive	2020/8/29	19	2020/8/29
400	69	Male	Nasal swab	Symptomatic	Positive	Positive	2020/9/2	18	2020/8/26
401	34	Female	Nasal swab	Symptomatic	Positive	Positive	2020/8/12	18	2020/8/8
402	69	Male	Nasal swab	Symptomatic	Positive	Positive	2020/8/31	29	2020/8/28
403	46	Male	Nasal swab	Symptomatic	Positive	Positive	2020/8/29	22	2020/8/24
404	30	Male	Nasal swab	Symptomatic	Positive	Positive	2020/8/7	24	2020/8/2
405	37	Male	Nasal swab	Symptomatic	Positive	Positive	2020/9/5	25	2020/8/31
406	55	Male	Nasal swab	Symptomatic	Positive	Positive	2020/8/15	22	2020/8/14
407	22	Male	Nasal swab	Symptomatic	Positive	Positive	2020/8/21	22	2020/8/15
408	60	Female	Nasal swab	Symptomatic	Positive	Positive	2020/8/17	19	2020/8/14
409	43	Male	Nasal swab	Symptomatic	Positive	Positive	2020/8/14	17	2020/8/10
410	58	Male	Nasal swab	Symptomatic	Positive	Positive	2020/9/5	32	2020/8/29
411	31	Female	Nasal swab	Symptomatic	Positive	Positive	2020/8/15	26	2020/8/10
412	22	Female	Nasal swab	Symptomatic	Positive	Positive	2020/8/12	27	2020/8/12
413	52	Female	Nasal swab	Symptomatic	Positive	Positive	2020/8/26	22	2020/8/21
414	40	Male	Nasal swab	Symptomatic	Positive	Positive	2020/8/2	24	2020/7/27
415	54	Female	Nasal swab	Symptomatic	Positive	Positive	2020/9/1	28	2020/8/27
416	51	Female	Nasal swab	Symptomatic	Positive	Positive	2020/8/20	31	2020/8/19
417	60	Male	Nasal swab	Symptomatic	Positive	Positive	2020/8/23	21	2020/8/17
418	40	Male	Nasal swab	Symptomatic	Positive	Positive	2020/8/1	20	2020/7/29
419	54	Male	Nasal swab	Symptomatic	Positive	Positive	2020/8/11	23	2020/8/4
420	51	Female	Nasal swab	Symptomatic	Positive	Positive	2020/8/20	25	2020/8/19
421	60	Female	Nasal swab	Symptomatic	Positive	Positive	2020/8/25	35	2020/8/21
422	40	Female	Nasal swab	Symptomatic	Positive	Positive	2020/8/3	23	2020/8/1
423	52	Male	Nasal swab	Symptomatic	Positive	Positive	2020/7/30	24	2020/7/28
424	47	Female	Nasal swab	Symptomatic	Positive	Positive	2020/8/22	22	2020/8/20